





50 pcs



20 pcs



Sichuan Ai Doctor Medical Technology Co.,Ltd
333 Yongke Road, Yongsheng town, Wenjiang District,
Chengdu City, Sichuan Province, China

EC REP: SUNGO Europe B.V.
Olympisch Stadion 24, 1076DE Amsterdam, Netherlands

Storage method: Masks shall be stored in a clean, ventilated and dry warehouse, with a temperature -10°C and a relative humidity ≤80%.
Precautions: It is recommended to replace it every 2-4 hours, once it is contaminated, it should be replaced as soon as possible;

SIZE: 17X9.5CM
Made in China

EN 14683 Type II



SIZE: 17X9.5CM
Made in China

EN 14683 Type II



HYGISUN

REF HS0401A



Medical Face Mask

Earloop.Blue

EN14683 Type II



> Retouradres Postbus 16114 2500 BC Den Haag

SUNGO Europe B.V.
T.a.v. de heer R. Luo
Olympisch Stadion 24
1076 DE Amsterdam

Datum: 29 mei 2020
Betreft: aanmelding medisch hulpmiddel klasse I

Geachte heer Luo,

Graag bevestig ik hierbij de ontvangst op 19 mei 2020 van de mededeling ex artikel 5 van het Besluit medische hulpmiddelen (BMH) dat bedrijf Sichuan Ai Doctor Medical Technology Co., Ltd. met Europees gemachtigde SUNGO Europe B.V. onderstaand medisch hulpmiddel, ingedeeld in risicoklasse I, aflevert. Het product is onder volgend kenmerk geregistreerd. Ik verzoek u om in alle verdere correspondentie betreffende dit product het bijbehorende kenmerk te vermelden.

**Surgical Mask (Non-sterile)
(geen merknaam) (NL-CA002-2020-51493)**

Toekomstige wijzigingen in bovengenoemde gegevens – waaronder een eventuele wijziging van de indeling in risicoklasse in verband met wijzigingen van Europese regelgeving inzake de classificatie van medische hulpmiddelen, en aan voortschrijdend wetenschappelijk inzicht (zie art.9, lid 3 van Europese Richtlijn 93/42/EEG) – dient u te zijner tijd mede te delen.

Volledigheidshalve wijs ik u erop dat het - ongeacht uw mededeling - verboden is een medisch hulpmiddel ter aflevering voorhanden te hebben, dan wel af te leveren indien niet aan de voor dat medisch hulpmiddel geldende regels gesteld bij of krachtens de Wet op de Medische Hulpmiddelen (WMH) wordt voldaan. Met name wijzen wij u op de Nederlandse-taaleis, de eisen voor het ter beschikking houden van de technische documentatie en de plicht tot het hebben van een Post Market Surveillance- en vigilantiesysteem.

Farmatec

Bezoekadres:
Hoftoren
Rijnstraat 50
2515 XP Den Haag
T 070 340 6161

<http://hulpmiddelen.farmatec.nl>

Inlichtingen bij:

R.A.C. Ori

medische_hulpmiddelen@
minvws.nl

Ons kenmerk:

CIBG-20202302

Bijlagen

-

Uw aanvraag

19 mei 2020


*Correspondentie uitsluitend
richten aan het retouradres met
vermelding van de datum en het
kenmerk van deze brief.*

Tevens wijs ik u er voor de goede orde nog op dat de registratie van uw mededeling betreffende de aflevering van het bovengenoemde product slechts een administratieve handeling betreft. Deze ontvangstbevestiging behelst dan ook geen besluit betreffende de kwalificatie van het desbetreffende product als medisch hulpmiddel in de zin van art. 1 WMH, noch betreffende de indeling in risicoklasse I.

**zer niet
gedefinieerd.**

De Minister voor Medische Zorg en Sport,
namens deze,

Afdelingshoofd
Farmatec

A handwritten signature in blue ink, consisting of a series of loops and a long horizontal stroke, positioned above the name Dr. M.J. van de Velde.

Dr. M.J. van de Velde

DECLARATION OF CONFORMITY

Regarding Medical Device Directive(93/42/EEC)
including Directive 2007/47/EC

Manufacturer: Sichuan Ai Doctor Medical Technology Co., Ltd
Address: No.333, Yongke Road, Yongsheng Town, Chengdu Cross-strait
Science and Technology Industrial Development Park, Wenjiang
District, Chengdu, Sichuan, China

EC Representative: SUNGO Europe B.V.
Address: Olympisch Stadion 24, 1076DE Amsterdam, Netherlands

Product Name: Surgical mask (Non-sterile)
Specification: Earloop Type 170*95MM

Classification: Class I (MDD, Annex IX), Rule 1(All non-invasive devices are in class I)
Conformity Assessment: Annex VII



We confirm our product can meet the requirement of Medical Device Directive(93/42/EEC)
and the following harmonized standards.

EN ISO 14971:2012

EN ISO 10993-5:2009

EN ISO 15223-1:2016

EN ISO 10993-10:2013

EN 1041:2013

*On behalf of SUNGO Europe office, I confirmed we are
EU REP of the company who issue this document.*

EN 14683:2019

EN ISO 10993-1:2009/AC:2010

Signature: 

Date: 2020年5月15日


Authorized Signature (S)





171021110579

检验检测报告

TEST REPORT

STFWT20201056

产品名称

Product Name

一次性使用日用口罩

委托单位

Trust Unit

四川艾医生医疗科技有限公司

生产单位

Manufacturer

四川艾医生医疗科技有限公司

检验检测类别

Test Category

委托送样检验



江苏省特种安全防护产品质量监督检验中心

JIANGSU QUALITY SUPERVISION AND INSPECTION CENTER FOR SPECIAL SAFETY PROTECTION PRODUCTS

检验检测报告
Test Report



防伪查询

STFWT20201056

共 3 页 第 1 页

Page 1 of 3

产品名称 Product Name	一次性使用日用口罩	规格型号 Specification Type	挂耳型 170mm*95mm
		商 标 Trademark	敏甄™
委托单位 Trust Unit	四川艾医生医疗科技有限公司	电 话 Tel	18215607950
生产单位 Manufacturer	四川艾医生医疗科技有限公司	样品等级 Sample Grade	KN90
样品数量 Sample Quantity	30 只	送样日期 Sample Receiving Date	2020-02-17
检验检测类别 Test Category	委托送样检验	批号/货号 Serial Number	MY20200201
样品状态 Samples Conditions	符合检测要求		
检验检测及判定依据 Document and Decide Accordance	YY/T0969-2013 《一次性使用医用口罩》 GB/T32610-2016 《日常防护型口罩技术规范》		
检验检测结论 Test Conclusion	样品经检验,符合 GB/T32610-2016 和 YY/T0969-2013 标准规定的要求。  签发日期: 2020-02-26 SignatiumDate		
备 注 Remarks	本报告检验结论仅对所检项目得出,不代表未经检验的项目或功能符合要求。 本报告仅对来样负责。		

批准:
Approver

顾海
甘

审核:
Examiner

吴高亮

主 检:
Major tester

蔡燕文

检 验 检 测 结 果

Testing Results

STFWT20201056

共 3 页 第 2 页
Page 2 of 3


序号 Serial	检验检测项目 Test Items	单位 Unit	技术要求 Requirement	检验检测结果 Results	单项评价 Individual Judgment	
1	外观	—	口罩外观应整洁，形状完好，表面不得有破损、污渍。	口罩外观整洁，形状完好，表面未有破损、污渍。	合格	
2	结构与尺寸	—	口罩佩戴好后，应能罩住佩戴者的鼻、口至下颌。应符合设计的尺寸，最大偏差应不超过±5%。	口罩佩戴好后，能罩住佩戴者的鼻、口至下颌。符合设计的尺寸	合格	
3	鼻夹	—	1、口罩上应配有鼻夹，鼻夹由可塑性材料制成。 2、鼻夹长度应不小于8.0cm。	1、口罩上配有鼻夹，鼻夹由可塑性材料制成。 2、鼻夹长度： 9.3cm、9.4cm、9.5cm。	合格	
4	口罩带	—	1、口罩带应戴取方便。 2、每根口罩带与口罩体连接点处的断裂强力应不小于10N。	1、口罩带戴取方便。 1 [#] : 试样断裂强力不小于10N 2 [#] : 试样断裂强力不小于10N 3 [#] : 试样断裂强力不小于10N	合格	
5	细菌过滤效率/% (BFE)	—	≥95	97.1	合格	
6	微生物指标	细菌菌落总数	CFU/g	≤100	4	合格
		大肠菌群	—	不得检出	未检出	
		金黄色葡萄球菌	—	不得检出	未检出	
		绿脓杆菌	—	不得检出	未检出	
		溶血性链球菌	—	不得检出	未检出	
		真菌	—	不得检出	未检出	

检 验 检 测 结 果
Testing Results

STFWT20201056

共 3 页 第 3 页
Page 3 of 3

样 品 图 片
测试样品



以下空白

注 意 事 项

- 1、检验检测报告无“检验检测报告专用章”或检验检测单位公章无效。
- 2、复制检验检测报告未重新加盖“检验检测报告专用章”或检验检测单位公章无效。
- 3、检验检测报告无主检、审核、批准人签字无效。
- 4、检验检测报告涂改无效。

Points For Attention

1. This Report Is Invalid If Without “The Text Report Special Seal” Or The Official Seal Of The Institute.
2. The Reproduction Is Invalid If Without Being Confirmed By “The Text Report Special Seal” Or The Official Seal Of The Institute.
3. This Report Is Invalid If Without Signature Of The Major Tester And The Examiner And The Approver.
4. This Report Is Invalid If In Any Form By Any Means Altered.

检验检测机构地址：江苏省泰州市高港区临港经济园临港大道166号

The Institute Add: Lingang Road 166, Lingang Economic Park , Gaogang, Taizhou.Jiangsu

检验检测机构监督电话：0523-86989901

The Institute Complain Tel:0523-86989901

检验检测机构业务电话：0523-86989959

The Institute Business Tel:0523-86989959

检验检测机构传真：0523-86989939

The Institute Fax:0523-86989939

检验检测机构邮编：225300

The Institute Post:225300

检验检测机构网址：www.jstfzx.com

The Institute Web:www.jstfzx.com

检验检测机构邮箱：1735889887@qq.com

The Institute E-mail:1735889887@qq.com

[HYGCEN AUSTRIA GMBH | WERKSGELÄNDE 28 | 5500 BISCHOFSHOFEN]




52146 Würselen
Deutschland

Akkreditierte
Prüfstelle nach
ÖNORM EN ISO 17025



Bischofshofen, 25.05.2020

Prüfbericht / test report B 24090

Labor-Nr. / <i>identification of the test laboratory:</i>	B 24090
Prüfprodukt / <i>test product:</i>	Maske HS0401A Hygisun
Auftraggeber / <i>ordered by:</i>	
Auftragsdatum / <i>date of order:</i>	2020-04-28
Materialeingang / <i>date of delivery:</i>	2020-05-04
Prüfzeitraum / <i>period of analysis:</i>	2020-05-11 bis / to 2020-05-25
Prüfbedingungen / <i>test conditions:</i>	Die Prüfung erfolgte im Anlieferungszustand. / <i>The test was done in the delivery state.</i>
Prüfauftrag / <i>test order:</i>	Medizinische Gesichtsmasken - Anforderungen und Prüfverfahren / <i>Medical face masks - Requirements and test methods</i> EN 14683:2019
Prüfmethoden / <i>test methods:</i>	SOP 13-002 Bakteriellen Filtrationsleistung (BFE) / <i>bacterial filtration efficacy (BFE)</i> EN 14683 Anhang / <i>annex B</i> SOP 13-001 Atmungs-Eignungsprüfung, differentialer Druck / <i>breathability test, differential pressure</i> EN 14683 Anhang / <i>annex C</i> SOP 07-014 Mikrobiologische Reinheit / <i>Determination of a population of micro-organisms</i> EN ISO 11737-1

Ergebnis der Prüfung der Filterwirksamkeit für Bakterien für Masken / test results of bacterial filtration efficacy of masks
EN 14683 / SOP 13-002

Information: 5.2.2 der EN 14683:2019+AC

Prüfprodukt / test product:	Maske HS0401A Hygisun
Prüfdatum / date of testing:	2020-05-12
Anzahl der Prüfkörper / number of samples:	5
Volumenfluss / volume flow:	28.3 l/min
Größe der Prüfkörper / sample size:	10cm x 10cm
Geprüfter Bereich des Prüfkörpers / sample area tested:	50 cm ²
Prüfkeim / test strain:	<i>Staphylococcus aureus</i> ATCC 6538
KBE der Ausgangskeimsuspension / cfu of test suspension:	5,72 x 10 ⁵ /ml
Inkubation / incubation:	48 h bei / at 36 ± 1 °C
Raumtemperatur / room temperature:	20 °C
Luftfeuchte / relative humidity:	40 %

Ergebnis der Prüfung der Filterwirksamkeit für Bakterien für Masken / test results of bacterial filtration efficacy of masks
EN 14683 / SOP 13-002

Gezählte KBE/Platte / counted cfu per plate

	Ebene 1 KBE/Platte level 1 cfu/plate	Ebene 2 KBE/Platte level 2 cfu/plate	Ebene 3 KBE/Platte level 3 cfu/plate	Ebene 4 KBE/Platte level 4 cfu/plate	Ebene 5 KBE/Platte level 5 cfu/plate	Ebene 6 KBE/Platte level 6 cfu/plate	KBE gesamt total cfu
PK1	217	298	321	357	342	298	1833
PK2	198	216	302	376	301	317	1710
NK	0	0	0	0	0	0	0

*Gezählte KBE/Platte nach Umrechnung mit „Positive hole conversion table“ /
 Counted cfu per plate after conversion with “positive hole conversion table”*

	Ebene 1 KBE/Platte level 1 cfu/plate	Ebene 2 KBE/Platte level 2 cfu/plate	Ebene 3 KBE/Platte level 3 cfu/plate	Ebene 4 KBE/Platte level 4 cfu/plate	Ebene 5 KBE/Platte level 5 cfu/plate	Ebene 6 KBE/Platte level 6 cfu/plate	KBE gesamt total cfu
PK1	217	298	649	892	772	547	3375
PK2	198	216	583	1125	559	629	3310
NK	0	0	0	0	0	0	0

Gezählte KBE/Platte / counted cfu per plate

Probe / sample	Ebene 1 KBE/Platte level 1 cfu/plate	Ebene 2 KBE/Platte level 2 cfu/plate	Ebene 3 KBE/Platte level 3 cfu/plate	Ebene 4 KBE/Platte level 4 cfu/plate	Ebene 5 KBE/Platte level 5 cfu/plate	Ebene 6 KBE/Platte level 6 cfu/plate	KBE gesamt total cfu
1	12	5	5	6	7	24	59
2	5	1	5	2	4	21	38
3	6	6	3	5	5	15	40
4	3	6	2	8	13	26	58
5	11	5	6	9	7	21	59

*Gezählte KBE/Platte nach Umrechnung mit „Positive hole conversion table“ /
 Counted cfu per plate after conversion with “positive hole conversion table”*

Probe / sample	Ebene 1 KBE/Platte level 1 cfu/plate	Ebene 2 KBE/Platte level 2 cfu/plate	Ebene 3 KBE/Platte level 3 cfu/plate	Ebene 4 KBE/Platte level 4 cfu/plate	Ebene 5 KBE/Platte level 5 cfu/plate	Ebene 6 KBE/Platte level 6 cfu/plate	KBE gesamt total cfu
1	12	5	5	6	7	25	60
2	5	1	5	2	4	22	39
3	6	6	3	9	5	15	44
4	3	6	2	8	13	27	59
5	11	5	6	9	7	22	60

Legende / legend:

- KBE / cfu = Kolonie bildende Einheiten / colony forming units
- PK = Positivkontrolle / positive control
- NK = Negativkontrolle / negative control

Bewertung der Filterwirksamkeit / rating of bacterial filtration efficacy
EN 14683 / SOP 13-002

Probe / sample	Filterwirksamkeit filtration efficacy [%]
1	98,20
2	98,83
3	98,68
4	98,23
5	98,20
Mittelwert mean value	98,52

Berechnungsformel / *calculation formula*: $B = \frac{(C - T)}{C \times 100}$

- C = Mittelwert der gesamten Plattenausählung für die beiden positiven Kontrollläufe
plate count average of both positive control runs
- T = gesamte Plattenausählung für das Prüfstück
total plate count of the sample

Ergebnis der Atmungs-Eignungsprüfung, differentialer Druck, in Übereinstimmung mit U. S. Militär-Spezifikation-MIL-M-36954 C (Luft-Austausch-druck) / breathability test result, differential pressure, in accordance with U.S. Military Specification MIL-M-36954 C (Air Exchange Pressure)
EN 14683 / SOP 13-001

Information: 5.2.3 der EN 14683:2019+AC

Prüfprodukt / test product: Maske HS0401A Hygisun
Prüfdatum / date of testing: 2020-05-18
Anzahl der Prüfkörper / number of samples: 5
Anzahl der Prüfungen pro Prüfkörper / number of tests per sample: 5
Größe der Prüfkörper / sample size: 10cm x 10cm
Geprüfter Bereich des Prüfkörpers / sample area tested: kreisförmig, Durchmesser 2,5 cm / circular, diameter 2.5 cm
Geprüfter Bereich des Prüfkörpers / tested area of the test sample: 4.9 cm²
Luftstrom / airflow: 8 l/min ± 0.2 l
Raumtemperatur / room temperature: 20 °C
Luftfeuchte / relative humidity: 48 %
Standardabweichung / standard deviation: 1,19

Probe / sample	Pos. 1 Pa	Pos. 2 Pa	Pos. 3 Pa	Pos. 4 Pa	Pos.5 Pa	Mittelwert / mean value Pa	ΔP [Pa/cm ²]
1	178,90	133,90	157,90	168,90	112,30	150,38	30,69
2	107,70	198,50	182,30	145,90	139,50	154,78	31,59
3	138,50	168,90	164,50	101,30	116,80	138,00	28,16
4	124,50	132,50	110,30	157,80	191,20	143,26	29,24
5	143,60	178,60	156,30	127,10	113,70	143,86	29,36
Mittelwert / mean value							29,81

Legende / legend:

Pa = Pascal

Berechnungsformel / calculation formula = Differentialdruck / differential pressure $\Delta P = \frac{\bar{x} m_2 - \bar{x} m_1}{4,9}$

Bestimmung der Population von Mikroorganismen auf Produkten, Mikrobiologische Reinheit / Determination of a population of micro-organisms on products EN ISO 11737-1 / SOP 07-014

Information: 5.2.5 der EN 14683:2019+AC

Ergebnis der Validierung des Ablösungsverfahrens mittels wiederholender Rückgewinnung / Result of the validation of the elution procedure using the method of repeated recovery

Prüfprodukt / test product: Maske HS0401A Hygisun
Prüfdatum / date of testing: 2020-05-11
Probengewicht / sample weight: 3.59 g
Ablösungsverfahren / dissolution procedure: 1 Maske wurde bei höchster Stufe im Stomacher 5 min mit Verdünnungslösung eluiert / 1 mask was processed 5 min in a stomacher by highest speed with dilution solution
Raumtemperatur / room temperature: 22.0 °C
Luftfeuchte / relative humidity: 41 %
Inkubation / incubation: Bebrütung des Membranfilters auf Blutagar, / Incubation of the membrane filter on blood agar, 48h bei / at 36 ±1°C

Anzahl Tests / number of tests	Volumen Eluierungsmittel / volume elution medium	KBE / Prüfkörper / cfu / test body
1a	20 ml	2
1b	20 ml	2
1c	20 ml	17
1d	20 ml	6

Berechnung der Ergebnisse / calculation of the results

Ablösung / dissolution in %: 7.41 %
 Korrekturfaktor / correction factor: 13.50

Legende / Legend:

VF = Verdünnungslösung / dilution solution (0.85% NaCl, 0.1% Trypton)
 KBE / cfu = Kolonie bildende Einheiten / colony forming units

$$\text{Ablösung / dissolution in \%} = \frac{\text{Anzahl der Mikroorganismen nach der 1. Eluierung / number of micro-organisms after the 1. elution}}{\text{Anzahl der Mikroorganismen von Eluierung 1 – 4 / number of micro-organisms after elution 1 – 4}} \times 100$$

$$\text{Korrekturfaktor / correction factor: } \frac{100}{\text{Ablösung / dissolution in \%}}$$

Bestimmung der Population von Mikroorganismen auf Produkten, Mikrobiologische Reinheit / Determination of a population of micro-organisms on products
EN ISO 11737-1 / SOP 07-014

Prüfprodukt / test product: Maske HS0401A Hygisun
Prüfdatum / date of testing: 2020-05-11
Probengewicht / sample weight: 3.59 g
Ablösungsverfahren / dissolution procedure: 1 Maske wurde bei höchster Stufe im Stomacher 5 min mit Verdünnungslösung eluiert / 1 mask was processed 5 min in a stomacher by highest speed with dilution solution
Raumtemperatur / room temperature: 22.0 °C
Luftfeuchte / relative humidity: 41 %
Inkubation / incubation: Bebrütung des Membranfilters auf Blutagar, / Incubation of the membrane filter on blood agar, 48h bei / at 36 ±1°C
Korrekturfaktor / correction factor*: 13.50

Nr. / No.	Nährboden / medium	Gesamtkeimzahl / KBE / PK total count cfu / PK	Mikrobiol. Differenzierung / microbiol. Differentiation	Gesamtkeimzahl x Korrekturfaktor total count x correction factor*	KBE / g cfu / g
Test 2	Blutagar / blood agar	4	-	54.00	15.04
Test 3	Blutagar / blood agar	2	-	27.00	7.52
Test 4	Blutagar / blood agar	6	-	81.00	22.56
Test 5	Blutagar / blood agar	1	-	13.50	3.76
Test 6	Blutagar / blood agar	3	-	40.50	11.28

Grenzwerte / Critical values:

Aerobe mesophile Keimzahl / aerobic mesophile germ number <1000 KBE / cfu
Hefen und Schimmelpilze / yeasts and moulds <100 KBE / cfu
Staphylococcus aureus <10 KBE / cfu
Sterptococcus <10 KBE / cfu
Pseudomonaden <10 KBE / cfu
Enterobacteriaceae <10 KBE / cfu

Ergebnis/ Result: 12,03 KBE/g

Legende / Legend:

VF = Verdünnungslösung / dilution solution (0.85% NaCl, 0.1% Trypton)
KBE / cfu = Kolonie bildende Einheiten / colony forming units
PK = Prüfkörper / test body
n = nicht zählbar / not countable

Schlussfolgerung / *conclusion:*

Die überprüfte Maske Maske HS0401A Hygisun entspricht den Vorgaben der EN 14683:2019, für Masken des Typs II.

The tested mask Maske HS0401A Hygisun fulfils the requirements of EN 14683:2019, for type II masks.

Archivierung / *Archiving:*

Eine Ausfertigung des Berichtes wird zusammen mit den Rohdaten im Archiv der HygCen Austria GmbH aufbewahrt. / *A copy of this report is kept together with the raw data in the archive of HygCen Austria GmbH.*

Hinweis / *Note:*

Der vorliegende Prüfbericht bezieht sich ausschließlich auf die dem Labor vorliegenden Prüfgegenstände. Jede auszugsweise Vervielfältigung bedarf der schriftlichen Genehmigung durch die HygCen Austria GmbH. / *The present test report refers exclusively to the test objects available to the laboratory. Any duplication in extracts requires the written permission of HygCen Austria GmbH.*



Prof. Dr. med. H.-P. Werner
Technischer Leiter / *technical manager*



Monika Feltgen
Stellvertreterin technischer Leiter / *vice technical manager*

Anhang zum Prüfbericht B 24090
attachment to test report B 24090



Abb. 1: Maske HS0401A Hygisun

Anhang / attachment
Erläuterung zum Prüfbericht B 24090
Comment to test report B 24090

1. Leistungsanforderungen für chirurgische Masken entsprechend / performance requirements for surgical masks EN 14683:2019

Prüfung / test	Typ / type I	Typ / type II	Typ / type IIR
Bakterielle Filterleistung / <i>bacterial filtration efficiency (BFE) %</i>	≥ 95	≥ 98	≥ 98
Druckdifferenz / <i>differential pressure Pa</i>	<40	<40	<60
Mikrobiologische Reinheit / <i>bioburden (g)</i>	≤ 30	≤ 30	≤ 30

2. Verfahren für die in-vitro Bestimmung der bakteriellen Filterleistung / method for in vitro determination of bacterial filtration efficiency (BFE)

Historie / history

Der Aufbau der Prüfung der Filterwirksamkeit für Bakterien für Masken wurde erstmals in der Militär-Spezifikation MIL-M-36954C „Mask, Surgical, Disposable“ aus dem Jahr 1975 beschrieben.

Seitdem wurde die Prüfung der Filterwirksamkeit in weiteren internationalen Normen umgesetzt:

von EDANA (European Disposables And Nonwovens Association) und INDA (Association of Nonwoven Fabrics Industry) in WSP 300.0 (05) „Standard Test Method for Nonwovens Bacterial Filtration Efficiency“,

von ASTM (American Society for Testing and Materials) in ASTM F 2101-07 „Standard Test Method for Evaluating the Bacterial Filtration Efficiency (BFE) of Medical Face Mask Materials, Using a Biological Aerosol of Staphylococcus aureus“ und

von CEN (Europäischen Normungskommission) in EN 14683 „Chirurgische Masken – Anforderungen und Prüfverfahren“.

The structure of the test method of the bacterial filtration efficiency for surgical masks was described for the first time in military specification MIL-M-36954C "Masks, Surgical, Disposable" from the year 1975. Since then the testing of the bacterial filtration efficiency was converted into further international standards:

by EDANA (European Disposables and Nonwovens Association) and INDA (Association OF Nonwoven Fabrics Industry) in WSP 300,0 (05) "standard test Method for Nonwovens Bacterial filtration Efficiency",

by ASTM (American Society for Testing and of material) in ASTM F 2101-07 "standard test Method for Evaluating the Bacterial filtration Efficiency (BFE) OF Medical Face MASK of material, Using A Biological aerosol OF Staphylococcus aureus"

and by the CEN (European standardization commission) in EN 14683 "surgical masks - requirements and testing methods".

Testprinzip / test principle

Eine Probe des Maskenmaterials wird zwischen ein sechsstufiges Kaskaden-Aufprallgerät (Andersen Sampler) und eine Aerosolkammer eingeklemmt. In die Aerosolkammer wird ein Aerosol von *Staphylococcus aureus* eingeführt und unter Vakuum durch das Maskenmaterial und das Aufprallgerät gezogen.

Die bakterielle Filterleistung der Maske wird durch die Anzahl der koloniebildenden Einheiten angegeben, die durch die Maske hindurchgehen, angegeben als Prozentsatz der im Belastungsmaterial vorliegenden koloniebildenden Einheiten.

A specimen of the mask material is clamped between a six-stage cascade impactor and an aerosol chamber. An aerosol of Staphylococcus aureus is introduced into the aerosol chamber and drawn through the mask material and the impactor under vacuum.

The bacterial filtration efficiency of the mask is given by the number of colony forming units passing through the surgical mask material expressed as a percentage of the number of colony forming units present in the challenge aerosol.

Auswertung der Prüfung / evaluation of the examination

Es wird eine Umrechnung der Keimzahlen auf den Ebenen 3-6 des sechsstufigen Kaskaden-Aufprallgerätes mit der „positive hole conversion table“ von Andersen A. A. (1958) durchgeführt.

The colony forming units on the levels 3-6 of the six-level cascade impact are converted with the "positive hole conversion table" described by Andersen A. A. (1958).

3. Verfahren zur Bestimmung der Atmungsaktivität (Druckdifferenz) / method for the determination of the breathability (differential pressure)

Testprinzip / test principle

In einem Versuchsaufbau wird die Druckdifferenz gemessen, indem Luft bei einem Luftstrom von 8l/min durch eine definierte Grundfläche des Materials gezogen wird.

Die Druckdifferenz wird mit einem Druckmessmodul gemessen. Gilmont-Instrumente-Flowmeter oder ein Flowmeter von vergleichbarer Präzision werden für die Messung des Luftstroms benutzt. Das Probenmaterial wird zwischen die Prüfflächen geklemmt, so dass das Probenmaterial quer zum Luftdurchfluss in der Strömung platziert ist. Die Vakuumpumpe wird eingeschaltet und die Durchflussrate der Luft wird am Flowmeter über ein Nadelventil auf 8l/min eingestellt. Durch das Druckmessmodul wird der Druck m_1 und m_2 gemessen und aufgezeichnet.

Dieses Verfahren wird an 5 unterschiedlichen Stellen der Materialien angewendet und der Mittelwert ermittelt.

In an experimental setup the differential pressure is measured, as air is pulled with an air flow by 8l/min through a defined surface area of the material.

The differential pressure is measured with a pressure measuring module. Gilmont-Instrumente-Flowmeter or a flowmeter of comparable precision are used for the measurement of the air flow. The sample material is wedged between the test surfaces, so that the sample material is placed transverse to the air flow. The vacuum pump is switched on and the flow rate of air is adjusted

at the flowmeter over a needle valve to 8l/min. The pressure m_1 and m_2 is measured and noted with the pressure measuring module.

This procedure will be performed at 5 different places of the materials and the mean value is determined.



SUBJECT Physical & Microbiological Test

TEST LOCATION TÜV SÜD China

TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai 201108, P.R. China

CLIENT NAME Sichuan Ai Doctor Medical Technology Co.,Ltd.

CLIENT ADDRESS No.333, Yongke Road, Yongsheng Town, Chengdu Cross-strait Science And
Technology Industrial Development Park, Wenjiang District, Chengdu, Sichuan,
China

TEST PERIOD 26-Apr-2020~08-May-2020

Prepared By

Bella Xu

(Bella Xu)
Report Drafter



Authorized By



(Leo Liu)
Authorized Signatory

Note: (1) General Terms & Conditions as mentioned overleaf. (2) The results relate only to the items tested.(3) The test report shall not be reproduced except in full without the written approval of the laboratory.(4) Without the agreement of the laboratory, the client is not authorized to use the test results for unapproved propaganda.

Chemical/Microbiology Laboratory:
TÜV SÜD Products Testing (Shanghai) Co.,
Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai
201108
P.R. China

Phone : +86 (21) 6037 6375
Fax : +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TÜV SÜD Certification and Testing
(China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China



TEST REPORT

Sample Description : Surgical Mask
Sample Quantity : 50 pieces
Lot Number/Batch Code : 20200401
Specification : Earloop Type
Size : /
Brand Name : /

Remark: The above information was provided by applicant.

Summary of Test Results

No.	Test Item	Test Method	Test Standard Type II R	Judgement
1	Bacterial Filtration Efficiency Test (BFE), %	EN 14683:2019+AC:2019(E) Annex B	≥ 98	Pass
2	Differential Pressure Test (Pa/cm ²)	EN 14683:2019+AC:2019(E) Annex C	< 60	Pass
3	Synthetic Blood Penetration Test (kPa)	ISO 22609:2004	≥ 16.0	Pass
4	Microbial Cleanliness Test (CFU/g)	EN 14683:2019+AC:2019(E) Annex D	≤ 30	Pass

Note: Pass = Meet customer requirements;

Fail = Fail customer requirements;

= No comment;

N.D. = Not detected.

Photo of Samples



Results

No.	Test Item	Test Result
1	Bacterial Filtration Efficiency (BFE) Test	Specimen 1#: 98.2%
		Specimen 2#: 98.0%
		Specimen 3#: 98.1%
		Specimen 4#: 98.1%
		Specimen 5#: 98.1%
2	Differential Pressure Test	26.6 Pa/cm ²
3	Synthetic Blood Penetration Test	Specimen 1#-32#: None seen
4	Microbial Cleanliness Test	Specimen 1#: <1 CFU/g
		Specimen 2#: 1 CFU/g
		Specimen 3#: <1 CFU/g
		Specimen 4#: <1 CFU/g
		Specimen 5#: <1 CFU/g

Bacterial Filtration Efficiency (BFE) Test

1. Purpose

For evaluating the bacterial filtration efficiency (BFE) of mask.

2. Sample description was given by client

Sample description : Surgical Mask
Specification : Earloop Type
Lot Number : 20200401
Sample Receiving Date : 2020-04-26

3. Test Method

EN 14683:2019+AC:2019(E) Annex B

4. Apparatus and materials

- 4.1 *Staphylococcus aureus* ATCC 6538.
- 4.2 Peptone water.
- 4.3 Tryptic Soy Broth(TSB).
- 4.4 Tryptic Soy Agar(TSA).
- 4.5 Bacterial filtration efficiency test apparatus.
- 4.6 Six-stage viable particle Anderson sampler.
- 4.7 Flow meters.

5. Test specimen

- 5.1 As requested by client, take a total of 5 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4 h at (21±5)°C and (85±5)% relative humidity.

6. Procedure

- 6.1 Preparation of the bacterial challenge: Dilute the culture in peptone water to achieve a concentration of approximately 5×10^5 CFU/mL.
- 6.2 Adjust the flow rate through the Anderson sampler to 28.3 L/min.
- 6.3 Deliver the challenge to the nebulizer using a syringe pump. Purge tubing and nebulizer of air bubbles.
- 6.4 Perform a positive control run without a test specimen to determine the number of viable aerosol particles being generated. The mean particle size (MPS) of the aerosol will also be calculated from the results of these positive control plates.
 - 6.4.1 Initiate the aerosol challenge by turning on the air pressure and pump connected to the nebulizer. Immediately begin sampling the aerosol using the Anderson sampler.
 - 6.4.2 Time the challenge suspension to be delivered to the nebulizer for 1 min.
 - 6.4.3 Time the air pressure and Anderson sampler to run for 2 min.
 - 6.4.4 At the conclusion of the positive control run, remove plates from the Anderson sampler.
- 6.5 Place new agar plates into Anderson sampler and clamp the test specimen into the top of the Anderson sampler, with the inside of the specimen facing towards the bacterial challenge (test area: 77cm^2).
- 6.6 Repeat the challenge procedure for each test specimen.
- 6.7 Repeat a positive control after completion of the sample set.
- 6.8 Perform a negative control run by collecting a 2 min sample of air from the aerosol chamber. No bacterial challenge should be pumped into the nebulizer during the collection of the negative control.
- 6.9 Incubate agar plates at $(37 \pm 2)^\circ\text{C}$ for (20 to 52) h.
- 6.10 Count each of the six-stage plates of the Anderson sampler.

7. Calculation

Total the count from each of the six plates for the test specimens and positive controls, as specified by the manufacture of Anderson sampler. The filtration efficiency percentages are calculated as follows:

$$\text{BFE} = (C - T) / C \times 100$$

T is the total plate count for the test specimen.

C is the mean of the total plate counts for the two positive controls.



8. Test results*

Stage Number \ P Value	Positive Control (A)	Positive Control (B)	Negative Control	Specimen 1#	Specimen 2#	Specimen 3#	Specimen 4#	Specimen 5#
1	53	98	0	0	0	0	1	1
2	112	136	0	2	1	1	2	1
3	130	165	0	1	1	2	6	1
4	213	298	0	3	1	6	7	6
5	1518	1518	0	26	29	27	27	29
6	434	475	0	16	19	13	7	12
Total (T), CFU	2460	2690	<1	48	51	49	50	50
Average (C), CFU	$2.6 \times 10^3 = (P_A + P_B) / 2$							
BFE, %				98.2	98.0	98.1	98.1	98.1
Requirements	≥ 98							
Remarks	<p><i>P</i> is the value of corresponding corrected particle counts as specified by the manufacturer of the cascade impactor. <i>T</i> is the total of <i>P</i> value for the test specimen. <i>C</i> is the mean of the total of <i>P</i> value of the two positive controls.</p>							

Differential pressure Test

1. Purpose

The purpose of the test was to measure the differential pressure of masks.

2. Sample description was given by client

Sample description : Surgical Mask
Specification : Earloop Type
Lot Number : 20200401
Sample Receiving Date : 2020-04-26

3. Test Method

EN 14683:2019+AC:2019(E) Annex C

4. Apparatus and materials

Differential pressure testing instrument

5. Test specimen

- 5.1 Test specimen are complete masks or shall be cut from masks. Each specimen shall be able to provide 5 different circular test areas of 2.5 cm in diameter.
5.2 Prior to testing, condition all test specimens for a minimum of 4 h at (21±5)°C and (85±5)% relative humidity.

6. Procedure

- 6.1 Without a specimen in place, the holder is closed and the differential manometer is zeroed. The pump is started and the flow of air adjusted to 8 L/min.
6.2 The pretreated specimen is placed across the orifice (total area 4.9cm², test area diameter 25mm) and clamped into place so as to minimize air leaks.
6.3 Due to the presence of an alignment system the tested area of the specimen should be perfectly in line and across the flow of air.
6.4 The differential pressure is read directly.
6.5 The procedure described in steps 6.1-6.4 is carried out on 5 different areas of the mask and readings averaged.

Results:

Specimen	Test Results* (Pa/cm ²)	Average (Pa/cm ²)	Requirements	Judgement
1#	27.4	26.6	< 60	Pass
2#	29.1			
3#	22.6			
4#	27.5			
5#	26.5			

Synthetic Blood Penetration Test

1. Purpose

For evaluation of resistance of masks to penetration by a fixed volume of synthetic blood at a high velocity.

2. Sample description was given by client

Sample description : Surgical Mask
Specification : Earloop Type
Lot Number : 20200401
Sample Receiving Date : 2020-04-26

3. Test Method

ISO 22609:2004

4. Apparatus and materials

- 4.1 Synthetic blood.
- 4.2 Tensiometer.
- 4.3 Synthetic blood penetration test apparatus;
- 4.4 Targeting plate.
- 4.5 Air pressure source.
- 4.6 Ruler.
- 4.7 Balance.
- 4.8 Controlled temperature and humidity chamber.

5. Test specimen

- 5.1 As requested by client, take a total of 32 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4h at $(21\pm 5)^{\circ}\text{C}$ and $(85\pm 5)\%$ relative humidity.

6. Procedure

- 6.1 Prepare the synthetic blood (40–44 mN/m) for the test.
- 6.2 Determine the density of the synthetic blood.
- 6.3 Fill the reservoir with new synthetic blood.
- 6.4 Position the test specimen 30.5 cm (12 in.) from the exit of the canula.
- 6.5 Set the reservoir pressure to the approximate pressure.
- 6.6 Place the targeting plate approximately 1 cm away from the mask.
- 6.7 Set the valve timer to 0.5 s. Collect and weigh the amount of fluid delivered (before the targeting hole).

- 6.8 Set the valve timer to 1.5 s. Collect and weigh the amount of fluid delivered (before the targeting hole).
- 6.9 Calculate the difference in weight of the two spurts. For a test fluid with a density of 1.003, Table 1 gives the target difference in weight plus lower and upper limits for a velocity range within 2% of the target.

Table 1 Target weight difference

Fluid Pressure (mmHg)	Weight difference for 1s difference in spurt duration (g)		
	Min.	Target	Max.
120	3.002	3.063	3.124

- 6.10 Adjust the reservoir pressure and repeat steps 6.7 to 6.9 until the weight difference is within the target range.
- 6.11 Record the weight difference for the spurts exiting the nozzle.
- 6.12 Record the pressure in the reservoir.
- 6.13 Set the valve time to 0.5 s. Collect and weigh the amount of fluid passing through the targeting hole.
- 6.14 Set the valve time to 1.5 s. Collect and weigh the amount of fluid passing through the targeting hole.
- 6.15 The difference in weight between the 0.5 s and 1.5 s spurts through the targeting plate shall be within +2 % ~ -5 % of the difference in weight from the nozzle.
- 6.16 If the differential weight is less than 95 % of the weight difference exiting the nozzle, check the aim of the stream to make sure it is passing cleanly through the targeting hole.
- 6.17 If the differential weight is more than 102 % of the weight difference exiting the nozzle, repeat the weight measurements exiting the nozzle (steps 6.7 to 6.11).
- 6.18 For standard synthetic blood, the timer duration can be estimated using the formula:
(ρ is the density of the test fluid.) $t = 0.5 + (2 \times \rho - g \text{ at } 0.5 \text{ s}) / (g \text{ at } 1.5 \text{ s} - g \text{ at } 0.5 \text{ s})$.
- 6.19 Record the timer setting to use as the starting point for subsequent testing.
- 6.20 Mount a test specimen on the specimen holding fixture. If the mask contains pleats, spread the pleats out when mounting the mask onto the fixture to present a single layer of material as the target area.
- 6.21 Squirt the synthetic blood onto the test specimen for the calculated time. Ensure that the synthetic blood hits the target area of mask.
- 6.22 Inspect the inside surface for synthetic blood penetration within 10 s of squirting the synthetic blood against the target area.
- 6.23 Report the results (none / penetration) for each test specimen at the test pressure.



Results:

Specimen	Test Results*	Requirements	Judgement
1#	None Seen	Pass Pressure at 16.0 kPa (120mmHg)	Pass
2#	None Seen		Pass
3#	None Seen		Pass
4#	None Seen		Pass
5#	None Seen		Pass
6#	None Seen		Pass
7#	None Seen		Pass
8#	None Seen		Pass
9#	None Seen		Pass
10#	None Seen		Pass
12#	None Seen		Pass
13#	None Seen		Pass
14#	None Seen		Pass
15#	None Seen		Pass
16#	None Seen		Pass
17#	None Seen		Pass
18#	None Seen		Pass
19#	None Seen		Pass
20#	None Seen		Pass
22#	None Seen		Pass
23#	None Seen		Pass
24#	None Seen		Pass
25#	None Seen		Pass
26#	None Seen		Pass
27#	None Seen		Pass
28#	None Seen		Pass
29#	None Seen		Pass
30#	None Seen		Pass
31#	None Seen		Pass
32#	None Seen		Pass

Microbial Cleanliness Test

1. Purpose

The purpose of the test was to measure microbial cleanliness of mask.

2. Sample description was given by client

Sample description : Surgical Mask
Specification : Earloop Type
Lot Number : 20200401
Sample Receiving Date : 2020-04-26

3. Test Method

According to EN ISO 11737-1:2018 to determine the microbial cleanliness of mask material, and refer to the procedure as described in EN 14683:2019+AC:2019(E) Annex D

4. Apparatus and materials

- 4.1 Orbital shaker.
- 4.2 0.45 um filter.
- 4.3 Tryptic Soy Agar (TSA).
- 4.4 Sabouraud Dextrose Ager (SDA) with chloramphenicol.
- 4.5 Formula of Extraction Liquid: 1g/L peptone, 5g/L NaCl and 2g/L Tween 20.
- 4.6 Extraction apparatus.

5. Test specimen

- 5.1 As requested by client, take a total of 5 mask samples.
- 5.2 Mask samples for testing are provided in the original primary packaging.
- 5.3 Condition at (18 to 26)°C and (45 to 65)% relative humidity during testing.

6. Procedure

- 6.1 Five test specimens are selected from the top, bottom and 3 randomly chosen marks.
- 6.2 The mask is aseptically removed from the packaging and placed in a sterile 500 mL bottle containing 300 mL of extraction liquid.
- 6.3 The bottle is laid down on an orbital shaker and shaken for 5 min at 250 rpm.
- 6.4 After extracting, 100mL of the extraction liquid is filtered through a 0.45 um filter and laid down on a TSA plate for the total viable aerobic microbial count. Another 100 mL aliquot of the same extraction liquid is filtered in the same way and the filter plated on SDA for fungi enumeration.
- 6.5 The plates are incubated for 3 days at 30°C and 7 days at (20 to 25)°C for TSA and SDA plates respectively.
- 6.6 Calculate the colonies of each agar plate.

7. Calculation

For each test specimen calculate the microbial cleanliness as follows by counting the total colonies of the TSA and SDA plates.

Results*:

Specimen	Colonies of the TSA Plate	Colonies of the SDA Plate	Microbial Cleanliness, (CFU/g)	Requirements	Judgement
1#	0	0	<1	EN14683:2019+AC:2019(E) Annex D EN ISO 11737-1:2018 ≤ 30 CFU/g	Pass
2#	0	1	1		
3#	0	0	<1		
4#	0	0	<1		
5#	0	0	<1		

Note:

- 1.*denotes this test was carried out by external laboratory assessed as competent.
2. This report is for internal use only such as internal scientific research ,education, quality control, product R&D.

-END OF THE TEST REPORT-

